


Using Fluid Walls for Single-Cell Cloning Provides Assurance in Monoclonality

SLAS Technology
2020, Vol. 25(3) 267–275
© 2019 Society for Laboratory
Automation and Screening
DOI: 10.1177/2472630319891135
journals.sagepub.com/home/jla


Cristian Soitu^{1*}, Cyril Deroy^{1*}, Alfonso A. Castrejón-Pita¹,
Peter R. Cook², and Edmond J. Walsh¹

Abstract

Single-cell isolation and cloning are essential steps in many applications, ranging from the production of biotherapeutics to stem cell therapy. Having confidence in monoclonality in such applications is essential from both research and commercial perspectives, for example, to ensure that data are of high quality and regulatory requirements are met. Consequently, several approaches have been developed to improve confidence in monoclonality. However, ensuring monoclonality using standard well plate formats remains challenging, primarily due to edge effects; the solid wall around a well can prevent a clear view of how many cells might be in a well. We describe a method that eliminates such edge effects: solid confining walls are replaced by transparent fluid ones, and standard low-cost optics can confirm monoclonality.

Keywords

fluid walls, single-cell cloning, optical clarity, monoclonality

Introduction

Single-cell isolation and cloning are critical steps in many high-value biological workflows, including the production of monoclonal antibodies,^{1–3} stem cell therapy,⁴ and gene editing.⁵ For these applications, it is of paramount importance that a selected colony derives from a single progenitor. The method of limiting dilution is commonly used to isolate single cells;^{6–10} a cell suspension is dispensed into wells of a microtiter plate using a distribution of less than one cell per well. Cell numbers per well follow Poisson's distribution. However, assurance that there is one, and only one, cell per well is challenging, so users often take cells from the resulting colony and perform a second round of cloning to increase confidence in monoclonality—adding significant time and cost to the workflow.

The challenge of ensuring monoclonality has been well documented by regulatory authorities,¹¹ who note that it is insufficient to show a single colony in a well during a cloning round. Instead, it should be demonstrated that only a single cell was originally plated in a well, ideally by imaging the entire well with sufficient resolution to identify the single progenitor. Identification of single cells in flat-bottomed well plates becomes complicated in the region where the well base meets the vertical wall. This junction produces an “edge effect,” which usually manifests itself as dark/blurred regions due to artifacts resulting from the effects of refractive index and curvature; this makes single-cell identification both challenging and time-consuming. Fluorescent

labeling is commonly used to determine if just one cell is present, but this brings additional cost and may adversely affect cell viability—and so may not be suitable for some workflows. Another difficulty when utilizing well plates is cell settling time; this can be reduced using a centrifuge, but this may also be harmful and laborious.

Many methods, more advanced and costly than limiting dilution, are routinely used or currently in development, to facilitate, automate, and accelerate cloning in standard microplate formats. The most established is fluorescence-activated cell sorting (FACS),^{12,13} which allows fluorescently labeled cells to be sorted rapidly and efficiently. However, cell labeling plus the high pressure and shear rates involved may adversely affect viability,^{14,15} especially of sensitive cells. Other methods include inkjet printing,^{16–18} manual cell picking, and the use of optical tweezers or laser

¹Department of Engineering Science, University of Oxford, Oxford, UK

²Sir William Dunn School of Pathology, University of Oxford, Oxford, UK

*These authors contributed equally to this work.

Received July 6, 2019, and in revised form Oct 24, 2019. Accepted for publication Nov 6, 2019.

Supplemental material is available online with this article.

Corresponding Author:

Edmond J. Walsh, Department of Engineering Science, University of Oxford, Oxford OX1 3PJ, UK.

Email: edmond.walsh@eng.ox.ac.uk

microdissection;¹⁵ some use label-free optical methods, and deposition into wells to give a higher fraction containing single cells than is achievable using limiting dilution—allowing one to “beat” Poisson’s limit. However, most provide confirmation that a single cell is present prior to deposition into a well, rather than after deposition. A wide range of microfluidic devices have also been developed for single-cell cloning,^{15,19,20} but have disadvantages associated with cell proliferation rate, colony retrieval, and nonstandard formats that are not easily integrated into existing workflows.

We now describe a simple microfluidic method that can provide confidence in monoclonality and is microplate compatible. Critically, rather than working around difficulties associated with edge effects, cell labeling, and spin-down requirements, we eliminate them by confining liquids within fluid walls rather than solid ones. In its simplest form, a sessile drop of media (perhaps containing a single cell) is placed on a standard polystyrene dish used for tissue culture, and some of the media is withdrawn by an automated aspirator. This leaves a relatively flat drop pinned to the dish by interfacial forces. The drop is bounded on the lower side by polystyrene and on the upper side by an air/media interface. This interface is optically transparent, and—after collecting an image using a standard microscope—one can obtain a clear view of whether the drop contains a single cell. Optical clarity can further be enhanced by overlaying the drop with an immiscible liquid that has a refractive index closer to that of the medium; we use a biocompatible fluorocarbon, FC40. Arrays of square droplets—which we call GRIDs—are easily made with chambers arranged in any chosen format and which can hold volumes from a microliter down to a few nanoliters.^{21–23} For proof of concept, we study GRIDs with 256 individual chambers sitting in a standard 60 mm tissue culture dish. In principle, the method can be incorporated into high-throughput workflows by increasing chamber density in a 60 mm dish (by 256-fold²²) or surface area (e.g., using a standard one-well microplate²²).

Materials and Methods

Theory

The maximum angle, μ_m , for which light rays are accepted by a microscope objective in air can be calculated with knowledge of the numerical aperture (NA):

$$\mu_m = \sin^{-1} NA$$

Rays with angles exceeding μ_m do not reach the image plane and so result in dark regions.²⁴ As rays pass through the curved liquid surface of a sessile drop, the change in refractive index results in refraction according to Snell’s law.

Consider a drop of water (refractive index $n = 1.33$) sitting on the bottom of a well in a polystyrene microplate ($n = 1.58$) in air ($n = 1$). If the drop radius is less than the capillary length ($\lambda_c < \sqrt{\frac{\gamma}{\Delta\rho g}}$, where λ_c corresponds to capillary

length, γ is surface tension, g is gravitational acceleration, and $\Delta\rho$ is the mass density difference of the fluids), then gravity has a negligible effect and the drop has the shape of a spherical cap. For a spherical cap (**Fig. 1A**, blue area), the maximum height (h), radius of curvature (R), volume (V_{cap}), footprint radius (a), and contact angle (θ) are related through²¹

$$V_{cap} = \frac{\pi (2 - 3\cos\theta + \cos^3\theta)}{3} \frac{a^3}{\sin^3\theta}$$

$$h = \frac{a}{\sin\theta} (1 - \cos\theta)$$

$$R = \frac{h^2 + a^2}{2h}$$

For a known volume and footprint area, the entire drop geometry can be evaluated. Then, the angle between a tangent at any point on the surface and the horizontal, α (**Fig. 1A,B**), is given by

$$\alpha = \sin^{-1} \frac{R_a}{R}$$

Using this angle, the trajectory of a light ray through a sessile drop can be determined using Snell’s law. **Figure 1B** illustrates paths of parallel rays through a drop on polystyrene as they pass through air/water interfaces (continuous lines) or FC40/water interfaces (dashed lines). If the rays are refracted such that their exit angle, μ , exceeds μ_m , then the image will appear dark in the regions shown in **Figure 1C**, where R_a coincides with $\mu = \mu_m$. In white regions, $\mu < \mu_m$, and in dark regions, $\mu > \mu_m$. When the drop is overlaid with an immiscible fluid such as the fluorocarbon, FC40, with $n = 1.29$, μ is lower than in air due to the higher refractive index (green paths in **Fig. 1B**).

Reagents and Equipment

All reagents and materials were purchased from Sigma-Aldrich (St. Louis, Missouri), unless otherwise specified. Matrigel (Corning, cat. 3542777) was used for **Figure 5A,iii**. FC40 was supplied by IotaSciences Ltd. (Oxford, UK). All GRIDs and sessile droplets were made on 60 mm tissue culture-treated cell culture dishes (Falcon, cat. 353002). Where indicated, droplets contain Allura Red dye.




Figure 1. Schematics illustrating drop parameters. **(A)** Geometry of a sessile drop. **(B)** Light paths through a sessile drop on a polystyrene substrate (for illustrative purposes only). Different incident angles at the drop surface (α_1 , α_m , α_2) result in different exit angles (μ_1 , μ_m , μ_2). **(C)** Refracted light enters the objective when $\mu < \mu_m$; when $\mu > \mu_m$, dark regions obstruct visibility of drop contents.

All fluidic arrangements were created and manipulated using a custom mini-printer (iotaSciences) and custom software. This printer consists of a three-axis traverse, which holds a Teflon rod (used to create fluid walls of FC40) and a dispensing needle connected to a syringe pump that is used to add or remove liquids to or from chambers in GRIDs and droplets.

Cell Culture

All cell lines were grown routinely in cell culture flasks at 5% CO₂ at 37 °C. Preliminary tests to demonstrate reliable single-cell identification in droplets (**Figs. 2 and 3**) used human embryonic kidney (HEK) cells cultured in Dulbecco's modified Eagle's medium (DMEM) + 10% fetal bovine serum (FBS; Gibco). Subsequent plating and cloning in GRIDs were performed using both HEK cells and mouse embryonic stem (ES) cells (EK.CCE cell line, derived from a single XY blastocyst-stage embryo of the 129/Sv//Ev strain²⁵). The latter were cultured in DMEM (Invitrogen) supplemented with 15% FBS, 1% penicillin/streptomycin (Invitrogen), 0.1 mM 2-mercaptoethanol, 1% glutamine (Invitrogen), 1% minimum essential media (MEM) nonessential amino acids (Gibco), 1 mM sodium pyruvate, and 1000 U/mL ESGRO leukemia inhibitory factor (LIF) on gelatin-coated plates.

Cells were plated by “scanning”²² in 1.9×1.9 mm chambers in GRIDs that contained 400 nL of medium, at densities of 0.1 and 1 cell/chamber. In assays assessing the effects of different feeding regimes (**Suppl. Figs. S2B and S3**), replenishment was done by removing 350 nL of old medium and adding the same volume of fresh medium, using the custom-built mini-printer.

Imaging Droplets

To identify single cells in standard 96-well plates, it is important that the entire region in the well where cells might have been deposited can be imaged clearly; in general, enhanced optical clarity lowers microscopy and labor costs. We will show that replacing the solid walls that confine cells with liquid ones improves optical clarity obtained using standard low-cost microscope objectives. This method was validated by placing eight 1 μ L sessile drops of cell media (DMEM + 10% FBS) on a standard polystyrene petri dish used for tissue culture. Fluid was extracted from seven drops using a syringe pump, leaving drops with 100–1000 nL (**Fig. 2**). All drops have the same footprint area, as FBS prevents the pinning line from receding,²⁶ and drops with the lowest volumes have the smallest contact angles. Drops were imaged for less than 10 s after formation to minimize evaporation using a Nikon D610 DSLR mounted




Figure 2. Images of drops of media with different volumes but identical footprints sitting in air on a polystyrene dish. Drops of 1000 nL were deposited and 0–900 nL removed. Footprints remain identical (diameters 1.68 mm) because media is firmly pinned to the dish as the receding contact angle is never reached (as that angle is sufficiently low). (A–C) Top (10× objective, NA = 0.25) and side views (taken with the FTA instrument) of 1000 nL drop, plus a cartoon illustrating the intensity along the dotted line across the diameter in A. (D–H) Top views of drops with smaller volumes (10× objective, NA = 0.25); reducing volume reduces curvature, so less of the chamber is obscured by the dark ring. (I) Top view of 1000 nL drop taken with a different objective (20×; NA = 0.75).

on an IX53 inverted microscope (Olympus), operating in bright-field mode, and fitted with a 10× objective (Olympus A10 PL; NA = 0.25). The method works in both bright-field and phase-contrast modes, although the former was exclusively used here. Contact angles, θ , were calculated using equations from the Theory section above from the measured footprint area (obtained using the calibration ruler in the microscope) and volume. They were also measured directly by the sessile drop method using the First Ten Angstroms (FTA) instrument and software. For the latter method, drops were formed by ejecting a 1 μ L drop using a needle (33G blunt NanoFil needle; World Precision Instruments) connected to a syringe pump (Harvard Ultra) through a Teflon tube. Drops were gently transferred to the surface of a square cut from the base of a Corning 60 mm suspension culture dish made from polystyrene, and then imaged from the side. The resultant equilibrium contact angles in air were $\sim 82^\circ$ and $\sim 80^\circ$ using the analytical and sessile drop methods, respectively.

Imaging GRIDs

Supplemental Figure S1A illustrates how GRIDs in standard tissue culture dishes are created, using a dispensing needle mounted on a three-axis traverse system. Cells are

plated in individual chambers in a GRID using a swift scanning motion. Due to the low volumes contained within these chambers (i.e., 400 nL) and the FC40 overlay, the rectangular footprints yield similar optical clarity to droplets in **Figure 3C,ii**.

For **Figures 4A, 4B, 5A,iii, and 5B** the same inverted microscope and objective was used to image each individual chamber.

Images in **Figures 4C, 5Ai, and 5Aii** were taken using an IncuCyte ZOOM, a commercial live-cell imaging system (Sartorius, Gottingen, Germany). Data illustrated in **Supplemental Figures S2 and S3** was collected using the IncuCyte and a second commercial imaging system, cytoSMART Omni (cytoSMART, Eindhoven, Netherlands). For these experiments, GRIDs were imaged immediately after plating single-cell suspensions. Images were analyzed to locate chambers containing single cells and cells in the chambers were fed using the regimes indicated (**Suppl. Figs. S2B and S3**); colonies were imaged after 8 days.

Results

Consider the image in **Figure 2A** of a 1000 nL drop of tissue culture medium sitting on a polystyrene petri dish in air. A black ring surrounds the gray central area. As the drop




Figure 3. Identifying cells in sessile drops. All drops have the same footprint area but contain 1000 or 400 nL \pm HEK cells. Images were collected immediately after plating, so cells had not yet adhered to the surface. Images were collected with a $10\times$ objective (NA = 0.25). (A–D) Sessile drops were prepared in 4 ways. (i) Illustrations of the experimental setup in each column. (ii) Images of drops of DMEM + 10% FBS under air or FC40. (iii) Similar drop to the one above but containing HEK cells. (iv) Digital magnification (0.4×0.4 mm) of the area shown above.

radius is less than the capillary length (see Materials and Methods), the drop's shape is that of a spherical cap (Fig. 2B). The radius, R_a , at the transition point between bright and dark regions was measured (Fig. 2C), and then light-ray paths and diffraction angles were calculated (Fig. 1B; Materials and Methods); some results are shown in the first row of Table 1.

Drops with smaller volumes but similar footprints were made by depositing 1000 nL and then removing up to 900 nL (Fig. 2D–H). Footprints remain unchanged on removal because the receding contact angle is sufficiently low thus is never reached; consequently, the media does not recede and remains firmly pinned to the dish,²² as illustrated in Movie 1. For all drops in Figure 2, a ray of light entering a drop vertically at R_a results in an average $\mu_m = 12.8^\circ$ (SD of 0.6°). The value of μ_m can be calculated using the NA of the objective and is $\mu_m = 14.5^\circ$. Considering the simplified assumption that parallel rays enter the drop, this agreement is satisfactory; moreover, the approximately constant value of μ_m for a range of drop volumes further supports the

validity of this analysis. The maximum refracted angle, μ_{max} , for all drops considered occurs at the pinning line of the 1000 nL drops and is 48° at a value of $\alpha = \theta = 82^\circ$. Imaging a drop with an objective with a higher numerical aperture (i.e., NA = 0.75, compared with 0.25) gave a dark ring with reduced width (Fig. 2I) and $\mu_m = 48.6^\circ$ (Table 1), as expected. Table 1 also shows that drops in panels G and H result in $\mu < 10^\circ$ everywhere, and dark regions disappear. The value of μ can be reduced further by overlaying drops with an immiscible fluid of refractive index closer to the one of the cell medium, like FC40 ($n = 1.29$); μ_m falls from an average of 12.8° in air to 1.3° in FC40 (Table 1). These results show that all volumes within a drop can be imaged, providing the drop volume is sufficiently small.

Imaging Cells in Sessile Drops

To evaluate whether individual cells can be seen clearly in such drops, drops (now containing HEK cells) with constant footprints but varying volumes were prepared on




Figure 4. Comparing optics in chambers with solid and fluid walls. **(A)** Images of colonies of HEK cells growing close to walls; the magnifications of areas indicated are shown below, where dotted lines mark clone limits. Left: The solid wall of a 96-well plate introduces an edge effect that partially obscures the view of the clone. Right: Fluid walls present no equivalent edge effect, and individual cells in the clone are seen clearly. **(B)** A single cell (left) and doublet (right) are easily identified immediately after plating, even though they lie close to fluid walls. **(C)** A gallery of five square chambers where the lengths of edges are reduced by half from 1.9 mm at the top, down to 110 μm , each containing one cell.




Figure 5. Cloning in chambers (1.9×1.9 mm) with fluid walls. **(A)** HEK cells. (i) Images of a whole chamber containing a single cell that develops into a colony over 8 days. The single cell seen at d0 moves slightly to the right in d4, prior to forming the colony. (ii) An arbitrary threshold (i.e., ≥ 16 cells derived from one progenitor) is used to define a colony; only the colony on the left meets this criterion. (iii) Fluidic chambers can be created on substrates \pm different coatings (each pair of images shows a view of the complete chamber on the left and zoom on the right). Coatings visibly affect the morphology of colonies, which ranges from compact (poly-L-lysine) to spread out (Matrigel). **(B)** Mouse ES cells grow as more compact colonies than HEK cells, imaged using a $4\times$ objective. The upper image shows a view of four chambers, each containing an 8-day-old colony derived from a single cell; zooms are shown below (dotted lines indicate colony limits). Edge effects are more pronounced here due to the addition of media for feeding, thereby increasing the GRIDs' contact angles (see Theory) resulting in the shaded regions around the edges. Single cells were identified prior to the addition of media.

Table 1. Geometric Parameters Calculated for Drops Like Those Shown in Figure 2 Assuming Their Shape Is Equivalent to the Cap of a Sphere.

Drop	Volume (μL)	R_0 (μm)	θ°	h (μm)	$\alpha^\circ @ R_0$	$\mu_m^\circ @ \alpha$ (air)	$\mu_m^\circ @ \alpha$ (FC40)
A	1	466	82	726	34	12.0	1.5
B	0.8	485	73	618	34	12.1	1.1
C	0.7	531	68	558	36	13.0	1.2
D	0.6	581	61	492	37	13.7	1.3
E	0.4	696	45	346	36	13.1	1.2
F	0.3	835	35	263	35	12.6	1.2
G ^a	0.2	835	24	177	24	8.2	0.8
H ^a	0.1	835	13	91	13	4.2	0.4
I ^a	1	835	82	726	82	48	10.9

^aCalculation based on $\alpha = \theta$ at the pinning line.

tissue culture-treated dishes and imaged immediately using a standard low-cost 10 \times objective (**Fig. 3**). As before, decreasing the volume and adding an FC40 overlay reduces the width of the dark ring, and the combination eliminates it completely; consequently, cells throughout the volume can be seen clearly. Of course, use of higher magnifications and numerical apertures further increases resolution. As 1000 nL drops have a substantial height (**Table 1**), some cells may not have settled on to the bottom when images are collected; however, they can still be detected, even though they may be out of focus (e.g., arrowed cells in **Fig. 3D,iv**). Of course, using drops with smaller volumes (and so heights) brings essentially all cells in the volume into sharp focus (**Fig. 3C,iv**). These results show how the slight edge effects associated with the use of fluid walls, and problems associated with the failure of cells to settle rapidly, are easily mitigated.

Comparing Optics in Conventional Wells with Their Equivalents Made with Fluid Walls

Although arrays of sessile drops under FC40 that have circular footprints are made easily,²¹ we now prefer to use alternatives with square footprints.²² This geometry is useful for creating a high number of chambers efficiently, while maximizing the surface area used. We call such arrays GRIDs. GRIDs are made by reshaping media and FC40 already sitting in a dish (**Suppl. Fig. S1**²²). First, the bottom of a standard petri dish is covered with a thin film of media (approximately 50 μL), a 1 mL FC40 overlay added to prevent evaporation, and a Teflon stylus (approximately 100 μm wide) lowered on to the bottom of the dish. As FC40 wets the stylus and polystyrene better than media, it runs down to the substrate. Moving the stylus sideways then “prints” a fluid wall of FC40 that is tightly attached to that substrate, held by interfacial forces. Drawing more lines then creates a GRID, in which neighboring chambers containing media are separated by liquid walls of FC40. GRIDs with hundreds to thousands of chambers containing

volumes down to a few nanoliters can be made, using the custom mini-printer. Here, we use GRIDs with 256 chambers made on 6 cm dishes in less than 90 s.²² Liquids can be added to, and removed from, individual chambers simply by pipetting through the liquid FC40 walls; then, within wide limits determined by advancing and receding contact angles, the fluid walls morph above an unchanging footprint. The walls also prove robust during handling,²² for example, a dish containing a GRID can be inverted (so essentially all FC40 drains out of the dish, leaving only a thin layer covering the wells and fluid walls—approximately a few microns high), but all chambers (and their aqueous contents) remain tightly attached to the dish (**Suppl. Fig. S1B**). Cells are deposited into chambers using a high-speed scanning motion²² performed by the mini-printer, in which a dispensing needle moves at constant speed and height above the chamber as it infuses a cell suspension. Each chamber receives the same volume (**Movie 2**) and—as heights are so small—cells settle quickly. Cells in GRIDs are then grown in conventional CO₂ incubators just like their counterparts in microplates, as FC40 is freely permeable to the vital gases, O₂ and CO₂. Previous work has shown that liquid FC40 walls effectively isolate reagents and cells in one chamber from others in the same GRID; this is largely due to the very low solubility of water in the fluorocarbon (i.e., <7 ppm by weight at room temperature).^{21–23} Consequently, after inoculating bacteria into every second chamber in a GRID, bacteria grow only in inoculated chambers as the rest remain sterile.²² While solid walls in a microplate provide an effective barrier between chambers, FC40 provides an additional sterile barrier over chambers. We now compare edge effects given by solid walls found in a conventional microplate with those given by the transparent and liquid ones in such GRIDs.

Figure 4A illustrates images of two colonies of adherent HEK cells, one growing near the solid wall of a microplate well, the other at a fluid wall in a GRID. The solid wall gives a significant edge effect (the dark annulus), and individual cells in the clone cannot be resolved even in the digital

zoom. In contrast, there are few edge effects near the fluid walls (which appear as light gray lines in the top panel), and individual cells throughout the colony are easily resolved in the zoom. It is even possible to distinguish between a single cell and two sitting side by side close to a fluid wall (**Fig. 4B**). Moreover, chamber dimensions are easily shrunk to the user's needs (by printing fluid walls closer together), without affecting cell visualization (**Fig. 4C**).

Cloning

As well as providing excellent optics for single-cell identification, GRIDs offer a suitable environment for cloning. After plating HEK cells in a GRID, an image scan across the GRID allows chambers containing single progenitors to be identified rapidly, and subsequently their growth into colonies can be tracked similarly (**Fig. 5A,i**). In addition to ensuring monoclonality, cloning efficiencies can be determined by identifying formed colonies soon after plating. For example, we used an arbitrary threshold of ≥ 16 cells derived from a single progenitor to call a clone the group of cells on the left in **Figure 5A,ii** (but not the group on the right). This call was made after 8 days, which contrasts with the 21 days we generally use with these cells in a well in a conventional microplate. The versatility of the platform also enables chambers to be made with different coatings, prior to seeding single cells (**Fig. 5A,iii**; the different coatings have complex effects on cell morphology, growth rate, and colony compaction²⁷). Additionally, the excellent optics and small chamber size even allow multiple colonies to be identified in a single image obtained with a standard 4 \times objective (**Fig. 5B**). Here, note the dark areas next to the fluid walls that vary in thickness (and are nonexistent at the corners). The chamber shape approaches the cap of a perfect sphere at the center, but this cap sits on a square footprint. Consequently, the contact angle along the pinning line varies from a minimum at each corner to a maximum halfway between each pair of adjacent corners; this underlies the variations in thickness of the dark areas (**Fig. 1**). As in the case of drops with circular footprints (**Fig. 3**), thickness at any specified point can be reduced by reducing the contact angle at that point (by withdrawing medium from the chamber). Notwithstanding these dark areas, it still generally remains possible to identify single cells and clones at any point in a chamber (as in the case of the clone in the chamber at the top right in **Fig. 5B**).

The process of dispensing cell suspensions into GRIDs is rapid. For example, it takes less than 90 s to plate cells into the 256 chambers in a GRID used here,²² and resulting cell distributions follow Poissonian statistics (**Suppl. Fig. S2A**). The same printer used to make GRIDs and plate cells can also replace spent media, and so feeding protocols are easily varied. For example, in **Supplemental Figure S2B** we show that frequent feeding has little impact on cloning

efficiency, suggesting that the 400 nL media initially present in a chamber is sufficient to support colony growth for 8 days. However, different feeding regimes can have slight effects on colony size (**Suppl. Fig. S3**). In addition, retrieval of colonies (after adding trypsin) and transfer into other GRIDs or conventional microtiter plates (by aspirating the detached cells) are also done using the printer. Retrieval efficiency was assessed by transferring colonies from fluid chambers to 96-well plates and 56 out of 57 samples reattached and grew.

Conclusion

Single-cell isolation and cloning is a core part of many workflows, and there is a growing need to ensure monoclonality. Unfortunately, this proves to be difficult and time-consuming when imaging wells in microplates, as solid walls obscure regions at the periphery, making it difficult—if not impossible—to ensure the well contains only one cell. As a result, methods for in-well identification of single cells usually require plates to be spun to settle cells, fluorescent tagging, or expensive optics. Here, we present a simple method to eliminate issues associated with edge effects, cell labeling, and settling time that uses standard bright-field or phase-contrast optics. Importantly, obscuring solid walls used conventionally to confine liquids are replaced by transparent fluid ones. Then, contact angles of the resulting drops are reduced so that the obscured region shrinks until contents of the whole drop become visible. Adding an overlay of the fluorocarbon, FC40, both prevents evaporation of small volumes and improves optical clarity, as its refractive index more closely matches that of the aqueous phase than air. Arrays of drops/chambers with fluid walls were created in which single cells and their resulting clones are identified easily. The method provides an in-well method for confirming monoclonality that is user-friendly, versatile, biocompatible, and easily adapted to high throughput.

Acknowledgments

We gratefully acknowledge the financial and technical support of IotaSciences. We thank the labs of Elizabeth Robertson (for mouse ES cells), David Greaves (for access to the IncuCyte live-cell imager), and cytoSMART (for allowing us to test the Omni).

Declaration of Conflicting Interests

The authors declared the following potential conflicts of interest with respect to the research, authorship, and/or publication of this article: Oxford University Innovation—the technology transfer company of the University of Oxford—has filed provisional patent applications on behalf of C.S., P.R.C., and E.J.W. partly based on this study. P.R.C. and E.J.W. each hold equity in, and obtain remuneration from, IotaSciences Ltd., a company that is exploiting this technology. IotaSciences Ltd. provides scholarships for C.S. and C.D.

Funding

The authors disclosed receipt of the following financial support for the research, authorship, and/or publication of this article: This work was supported by IotaSciences Ltd. (C.S., C.D.), the European Commission for a 7th Framework Marie Curie Grant contract no. 333848 (E.J.W.), a Royal Society University Research Fellowship (A.A.C.-P.), the Impact Acceleration Account of the Biotechnology and Biological Sciences Research Council (E.J.W., P.R.C.), the Medical Research Council for a Confidence in Concept award MC_PC_15029 (E.J.W., P.R.C.) and award MR/K010867/1 (P.R.C.).

References

1. Le, H.; Vishwanathan, N.; Jacob, N. M.; et al. Cell Line Development for Biomanufacturing Processes: Recent Advances and an Outlook. *Biotechnol. Lett.* **2015**, *37*, 1553–1564.
2. Sliwkowski, M. X.; Mellman, I. Antibody Therapeutics in Cancer. *Science* **2013**, *341*, 1192–1198.
3. Corsiero, E.; Bombardieri, M.; Carlotti, E.; et al. Single Cell Cloning and Recombinant Monoclonal Antibodies Generation from RA Synovial B Cells Reveal Frequent Targeting of Citrullinated Histones of NETs. *Ann. Rheum. Dis.* **2016**, *75*, 1866–1875.
4. Kropp, C.; Massai, D.; Zweigerdt, R. Progress and Challenges in Large-Scale Expansion of Human Pluripotent Stem Cells. *Process Biochem.* **2017**, *59*, 244–254.
5. Byrne, S. M.; Mali, P.; Church, G. M. Genome Editing in Human Stem Cells. *Methods Enzymol.* **2014**, *546*, 119–138.
6. Fuller, S. A.; Takahashi, M.; Hurrell, J. G. Cloning of Hybridoma Cell Lines by Limiting Dilution. *Curr. Protoc. Mol. Biol.* **1988**, *1*, 11.8.1–11.8.2.
7. Morley, A. A.; Trainor, K. J.; Seshadri, R. S. Cloning of Human Lymphocytes Using Limiting Dilution. *Exp. Hematol.* **1983**, *11*, 418–424.
8. Quiroz, J.; Tsao, Y. S. Statistical Analysis of Data from Limiting Dilution Cloning to Assess Monoclonality in Generating Manufacturing Cell Lines. *Biotechnol. Prog.* **2016**, *32*, 1061–1068.
9. Staszewski, R. Murphy's Law of Limiting Dilution Cloning. *Stat. Med.* **1990**, *9*, 457–461.
10. Trainor, K. J.; Morley, A. A. Cloning of Lymphocytes from Whole Blood by Limiting Dilution. *J. Immunol. Methods* **1983**, *65*, 369–372.
11. Kennett, S. In Establishing Clonal Cell Lines—A Regulatory Perspective. Presented at WCBP 2014: 18th Symposium on the Interface of Regulatory and Analytical Sciences for Biotechnology Health Products, Washington, DC, January 28–30, 2014.
12. Wang, J.; Min, Z.; Jin, M.; et al. Protocol for Single Cell Isolation by Flow Cytometry. *Single Cell Sequencing and Systems Immunology*; Springer Netherlands: Dordrecht, 2015; pp 155–163.
13. Herzenberg, L. A.; Parks, D.; Sahaf, B.; et al. The History and Future of the Fluorescence Activated Cell Sorter and Flow Cytometry: A View from Stanford. *Clinical Chemistry* **2002**, *48*, 1819–1827.
14. Mollet, M.; Godoy-Silva, R.; Berdugo, C.; et al. Computer Simulations of the Energy Dissipation Rate in a Fluorescence-Activated Cell Sorter: Implications to Cells. *Biotechnol. Bioeng.* **2008**, *100*, 260–272.
15. Gross, A.; Schöndube, J.; Zimmermann, S.; et al. Technologies for Single-Cell Isolation. *Int. J. Mol. Sci.* **2015**, *16*, 16897–16919.
16. Gross, A.; Schöndube, J.; Niekrawitz, S.; et al. Single-Cell Printer: Automated, On Demand, and Label Free. *J. Lab. Autom.* **2013**, *18*, 504–518.
17. Liberski, A. R.; Delaney, J. T.; Schubert, U. S. “One Cell–One Well”: A New Approach to Inkjet Printing Single Cell Microarrays. *ACS Comb. Sci.* **2011**, *13*, 190–195.
18. Shuichi, Y.; Akira, U.; Yoshitake, A.; et al. Cell Patterning through Inkjet Printing of One Cell per Droplet. *Biofabrication* **2012**, *4*, 045005.
19. Chabert, M.; Viovy, J. L. Microfluidic High-Throughput Encapsulation and Hydrodynamic Self-Sorting of Single Cells. *Proc. Natl. Acad. Sci. U.S.A.* **2008**, *105*, 3191–3196.
20. Zhang, Q.; Wang, T.; Zhou, Q.; et al. Development of a Facile Droplet-Based Single-Cell Isolation Platform for Cultivation and Genomic Analysis in Microorganisms. *Sci. Rep.* **2017**, *7*, 41192.
21. Walsh, E. J.; Feuerborn, A.; Wheeler, J. H. R.; et al. Microfluidics with Fluid Walls. *Nature Commun.* **2017**, *8*, 816.
22. Soitu, C.; Feuerborn, A.; Tan, A. N.; et al. Microfluidic Chambers Using Fluid Walls for Cell Biology. *Proc. Natl. Acad. Sci. U.S.A.* **2018**, *115*, E5926–E5933.
23. Soitu, C.; Feuerborn, A.; Deroy, C.; et al. Raising Fluid Walls around Living Cells. *Science Adv.* **2019**, *5*, eaav8002.
24. Campbell, J. M.; Christenson, H. K. Dynamic Measurement of Low Contact Angles by Optical Microscopy. *ACS Appl. Mater. Interfaces* **2018**, *10*, 16893–16900.
25. Robertson, E.; Bradley, A.; Kuehn, M.; et al. Germ-Line Transmission of Genes Introduced into Cultured Pluripotential Cells by Retroviral Vector. *Nature* **1986**, *323*, 445–448.
26. Prastowo, A.; Cook, P. R.; Walsh, E. J. Strong Contact Line Pinning of Sessile Drops Induced by Fetal Bovine Serum and a Serum-Free Replacement. **2019**, manuscript in preparation.
27. Liberio, M. S.; Sadowski, M. C.; Soekmadji, C.; et al. Differential Effects of Tissue Culture Coating Substrates on Prostate Cancer Cell Adherence, Morphology and Behavior. *PLoS One* **2014**, *9*, e112122.

Using fluid walls for single-cell cloning provides assurance in monoclonality

Cristian Soitu^{1†}, Cyril Deroy^{1†}, Alfonso A. Castrejón-Pita¹, Peter R Cook², Edmond J.

Walsh^{1*}

¹ Department of Engineering Science, University of Oxford, Oxford OX1 3PJ, UK,

² Sir William Dunn School of Pathology, University of Oxford, Oxford OX1 3RE, UK.

*Corresponding author, edmond.walsh@eng.ox.ac.uk

[†]Authors contributed equally

Keywords: fluid walls, single cell cloning, optical clarity, monoclonality

Supplementary information




Figure S1. Making GRIDs, and some demonstrations of their stability. (A) Steps in making a GRID. 1 mL of cell-culture media (DMEM) containing 10% FBS is added to a virgin Petri dish, and subsequently most (~0.95 mL) removed by pipetting. This leaves a very thin film of

media that has wetted the surface of the dish (approximately 50 μ L). 1 mL of FC40 is then added on top of the media to limit evaporation. A stylus made of polytetrafluoroethylene (PTFE, Teflon) is lowered on to the surface of the dish. As FC40 wets Teflon and polystyrene better than water, the fluorocarbon becomes pinned to the substrate around the tip of the stylus. The stylus is then dragged along the surface in the pattern shown; it pushes media aside, to leave a fluid wall of FC40 pinned to the polystyrene substrate. In the case shown, one central chamber isolated by fluid FC40 walls from surrounding media is created. “Printing” more lines (not shown), can then subdivide the original body of media into the wanted number of chambers. These chambers are used like the wells of a microplate, except that liquids are pipetted into them, and removed from them, through FC40 instead of air. **(B)** Stability of two GRIDs each containing 256 square chambers (1.9 x 1.9 mm) in a 6 cm dish. GRIDs are made as in **(A)** using tissue-culture media. Each chamber initially contains ~200 nL, and—as media volumes are too small to be seen—200 nL red dye is added to each chamber to aid visualization. **(i)** Two GRIDs after manufacture. FC40 is transparent, and so invisible. **(ii)** The left-hand dish is tilted by 90°; most FC40 has spilt out, but a small residue can be seen at the bottom corner of the dish—chambers remain firmly pinned by interfacial forces to the dish. **(iii)** The left-hand dish has now been almost completely flipped over; chambers still remain pinned to the dish, and FC40 drains out almost entirely. This illustrates the dominance of interfacial forces at the microscale over gravitational ones, and how firmly chambers are attached to the dish.




Figure S2. Cloning cells by limited dilution yields the expected Poissonian statistics, and efficiencies. A suspension of HEK cells, or mouse ESCs, are deposited by scanning²³ (~400 nL in each chamber in a GRID with 256 chambers of 1.9 x 1.9 mm in a 6 cm dish), and the number of cells per chamber immediately determined by microscopy using a 4X objective (as in Fig. 4B,C). **(A)** The process of dispensing a single-cell suspension of HEK cells in GRIDs does not introduce any bias in the distribution. Cells were seeded at densities, λ , of 0.1 and 1 cell/chamber. Distributions of numbers of cells in each chamber follow Poisson's distribution. **(B)** Cloning efficiencies of the two cell types assessed using three different feeding regimes. Cells are plated and imaged as in (A), grown, and some fed by removing 350 nL spent medium and replacing it with 350 nL fresh medium. Chambers are reimaged after 8 days of growth, and the number of colonies with ≥ 16 cells derived from a single progenitor are recorded. The resulting cloning efficiencies are in the expected ranges.




Figure S3. Size of colonies. Cells were cloned and fed as described in **Figure S2**, and areas occupied by clones scored in the images collected after 8 days. Areas occupied by the clones are as expected, and the different feeding regimes have marginal effects on colony size. This indicates that the 400 nL media in a chamber is nevertheless sufficient to support early clonal growth.

Movie 1. Withdrawing fluid from a droplet. A withdrawing needle is withdrawing fluid from a sessile drop deposited under FC40. As the fluid is extracted, the pinning line remains fixed and does not recede.

Movie 2. Plating cells using a scanning motion. A needle infusing cell suspension at constant flowrate traverses at constant height above equally-spaced fluidic chambers, such that each one receives the same volume.